



UK DECLARATION OF CONFORMITY

Issued in accordance with The Product Safety and Metrology etc. (Amendment) Regulations 2021.

We,

**Moore Industries
16650 Schoenborn Street
North Hills, CA 91343**

as the manufacturer, declare under our sole responsibility that the product(s),

330R2 Process and Temperature Meter

to which this declaration relates, is in conformity with relevant statutory requirements:

Electromagnetic Compatibility Regulations 2016

Electrical Equipment (Safety) Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment (Amendment) Regulations 2022

This conformity is based on compliance with the application of relevant designated standards or technical specifications.

Standards:

EN 55011:1998
EN 61000-6-2:2001
EN 61010-1:1995
EN 61326-1:2006
EN 63000:2018

The standards EN 55011:1998, EN 61000-6-2:2001, EN 61010-1:1995, and EN 61326-1:2006 are no longer designated. The requirements of these standards have been checked against the designated standards EN 55011:2016+A2:2021, EN 61000-6-2:2019, EN 61010-1:2010+A1:2019, and EN 61326-1:2021 and there were no major technical changes affecting the latest technical knowledge for the products listed above.

Product Markings:



UKCA mark was first affixed to the product in 2022

Signed for and on behalf of Moore Industries:

Name: Tigran Grigorian

Company: Moore Industries International, Inc.

Title: QA Manager

Date: February 5, 2024

Document No: UK DoC PD765MI {020524}